

HiMedia Laboratories Private Limited

C-40, Road No.21Y, MIDC, Wagle Industrial Area,
Thane(W) - 400604 , Website : www.himedialabs.com,
Email : info@himedialabs.com

Certificate of Analysis, Quality and Conformity

Material Code : DD008	Material Name : ONPG (50 discs / vI)	Lot No : 0000629324
Report No.: 40001438610	Date of Release & Report : 2024-02-07	Expiry Date : 2025-02

Appearance

Filter paper discs of 6 mm diameter bearing letters "On" in continuous printing style.

Cultural response

ONPG reaction observed in 0.85% sodium chloride solution of following culture containing ONPG (DD008) disc after an incubation of upto 4 hours at 35-37°C.

Organism	ONPG
Cultural response	
<i>Citrobacter freundii</i> ATCC 8090	Positive reaction: yellow colour
<i>Enterobacter aerogenes</i> ATCC 13048	Positive reaction: yellow colour
<i>Escherichia coli</i> ATCC 25922	Positive reaction: yellow colour
<i>Salmonella Choleraesuis</i> ATCC 12011	Positive reaction: yellow colour
<i>Proteus vulgaris</i> ATCC 13315	Negative reaction: no colour change
<i>Salmonella Typhimurium</i> ATCC 14028	Negative reaction: no colour change

- . ATCC is a registered trade mark of the American Type Culture Collection
- . NCTC and National Collection of Type Culture are registered trade mark of the Health Protection Agency

Control Media :

- . For Bacteria : Soyabean Casein Digest Agar / Columbia Blood Agar base enriched with 5% v/v Sheep/Horse blood.
- . For Yeast & Mold : Sabouraud Dextrose Agar.

- . All ISO 11133 : 2014/Amd.1:2018(E) control strains are included in the Quality parameter
- . HiMedia Laboratories Pvt Ltd is Certified for ISO 9001:2015, ISO 13485:2016 and WHO GMP

- . The Quality Assurance Parameters are as per the guidelines specified in CLSI (NCCLS) document M22-A3 wherever applicable.

Storage & Shelf Life

Store between 2-8°C. Use before expiry date on the label.

STATUS OF THE MATERIAL : APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current HiMedia manual or product sheets. The results reported were obtained at the time of release.

This document has been produced electronically and is valid


Gowri V

Microbiologist/Sr.Executive
Microbiologist


Ujwala M. Kokate

Asst./Dy/QC Manager


Dr. Santosh Kaul

Dy/QA Manager